

Measurement of Selected Halogenated Contaminants in Human Serum and Milk using GCxGC-IDTOFMS

Jef Focant¹, Andreas Sjödin², Wayman Turner², Don Patterson, Jr.²

¹University of Liège

²Centers for Disease Control and Prevention, Atlanta, GA

INTRODUCTION

A new method using comprehensive two-dimensional gas chromatography and isotope dilution time-of-flight mass spectrometry (GCxGC-IDTOFMS) for the simultaneous measurement of selected polychlorinated biphenyls (PCBs), organochlorine pesticides (OCPs), and brominated flame retardants (BFRs) is presented. Conversely to reference methods based on classical GC-MS, a single injection of the extract containing all species of interest conducts to accurate identification and quantification. GCxGC ensures the chromatographic separation of most compounds and TOFMS allows mass spectral deconvolution of co-eluting compounds as well as the use of ¹³C-labeled internal standards for quantification. Isotope ratio measurements of the most intense ions for both natives and labels ensure the required specificity. Potentially interfering matrix compounds are usually kept away from the compounds to be measured in the chromatographic space. The use of this new method with automated sample preparation procedures developed at the Centers for Disease Control and Prevention (CDC) for the analysis of human serum and milk [1,2] compared favorably to conventional isotope-dilution one-dimensional gas chromatography-high resolution mass spectrometry (GC-IDHRMS) for the different sample pools that were tested

EXPERIMENTAL

Standards, Chemicals and Supplies. All standard solutions were purchased from Cambridge Isotope Laboratories (Andover, MS, USA). The EC-5022 PCB 10-points calibration standard solution contained a mixture of 38 native PCBs spanning the concentration range 0.5 to 1000 pg/μL and ¹³C₁₂-labeled PCBs at a concentration of 75 pg/μL in nonane. The EC-5087 ¹³C₁₂-labeled PCB internal standard spiking solution contained ¹³C₁₂-labeled PCBs at a concentration of 7.5 pg/μL in methanol. The EO-5159 BFR 10-points calibration standard solution contained a mixture of native, spanning the concentration range 0.2 to 2000 pg/μL, and ¹³C₁₂-labeled analytes, at a concentration of 75 pg/μL. The EO-5158 ¹³C₁₂-labeled BFR internal standard spiking solution ¹³C₁₂-labeled compounds at a concentration of 7.5 pg/μL in methanol. The ES-5019 persistent OCP 8-points calibration standard solution contained a mixture of native and ¹³C-labeled analytes. Natives were in the concentration range of 5 to 1000 pg/μL, and ¹³C_n-labeled at a concentration of 100 or 250 pg/μL. The ES-5177 ¹³C_n-labeled OCP internal standard spiking solution contained (10 pg/μL in methanol) was used. We prepared the multi-analyte calibration solution (59 native compounds) by combining equal volumes of the EC-5022, EO-5159, and ES-5019 solutions. All details concerning consumables, as well as the glassware pre-cleaning are available elsewhere [1,2].

Samples. Human serum samples corresponded to a pool collected from 15 individuals in 2002 in 3 U.S. cities (Philadelphia, PA; Memphis, TN; Miami, FL) and obtained from the Memphis, TN Interstate blood bank. A mixture of water (3.5 mL) and calf serum (0.5 mL) (Bibco BRL; Grand Island, NY) was used as serum blank. Three human milk pools were analyzed. Pool A was obtained from the Mothers' milk bank (Denver, CO) and was a composite pool of 2 individuals collected in 2002, pool B and C both corresponded to 10 specimens collected in 2003 in California and in North-Carolina, respectively. A 10-fold water diluted bovine milk obtained in a local supermarket was used as method blank samples.

Extraction and Cleanup. A semi-automated extraction and cleanup method recently developed at CDC for the measurement of the PCBs, OCPs, and BFRs in human serum and milk has been used and is described in details elsewhere [1-3].

GC-IDHRMS analysis. They were performed on a MAT95XP instrument (ThermoFinnigan MAT, Bremen, Germany) interfaced with a 6890N gas chromatograph (Agilent Technologies, Atlanta, GA) fitted with a 15 m x 0.25 mm i.d. x 0.10 μ m film thickness DB-5HT capillary column (J&W Scientific, Folsom, CA). Details of the GC-IDHRMS analyses are given elsewhere [1].

GCxGC-IDTOFMS analysis. The GCxGC-TOFMS instrument was the Pegasus 4D (Leco Corp., St Joseph, MI). This system was based on a non-moving quadrupole jet dual stage modulator made of two cold nitrogen jets and two pulsed hot air jets responsible for trapping and refocusing of compounds eluting from the ^1D column. This modulator was mounted in an Agilent 6890 GC oven and liquid nitrogen was used to create the cold jets. Details regarding the system have been reported elsewhere [4,5]. The GC inlet temperature was 280°C for 1.2 μ L splitless injections. Carrier gas was helium and a constant flow of 0.8 mL/min was used. The GC column set used was made of the combination of a 15 m x 0.25 mm i.d. DB-1 100% dimethylpolysiloxane (J&W Scientific) with a film thickness of 0.25 μ m as ^1D and a 1.2 m x 0.10 mm i.d. high temperature HT-8 (8% Phenyl)-polycarborane-siloxane (SGE, Austin, TX) with a film thickness of 0.10 μ m as ^2D . Deactivated universal presstight connectors (Restek Corp., Bellefonte, PA) were used for connecting the capillary columns. During chromatographic separation, the primary GC oven was programmed as follows: 90°C for 1 min, then to 150°C at 10°C/min, and finally to 300°C at 1°C/min. The ^2D column was 40°C higher than the primary oven and operated in the iso-ramping mode. The temperature of the modulator had an offset of 60°C compared to the temperature of the primary GC oven. The modulator period was 3 s (0.33 Hz modulation frequency) with a hot pulse duration set at 700 ms and a cool time between stages of 800 ms. The MS transfer line was at 250°C. The source temperature was 250°C with a filament bias voltage of -70 V. The data acquisition rate was 60 spectra/s for a collected mass range of 100 to 750 m/z . Table 1 lists the masses selected for quantification. The detector voltage was 1800 V. Data collection and processing were achieved using the 2.10 version of Leco ChromaTOFTM software provided with the instrument. Peak apex finding was performed automatically and further manually corrected when required. The combination of slices corresponding to a compound was performed by automatically comparing the mass spectra under pre-established match criteria. Spectral searching was performed using the NIST library available with the software, as well as through the custom built ^{13}C -labeled compound library.

RESULTS AND DISCUSSION

The last compound (BDE-154) eluted after 45 min (analytical speed of 1.3 analytes /min). Most of the compounds were chromatographically resolved from each other. Heptachlor epoxide and Tetra-CB-74, as well as Hepta-CB-189 and Octa-CB-196 were two co-eluting couples that required the use of the deconvolution capability to be resolved. Figure 1 shows chromatograms of a multi-analyte standard solution and a real sample extract. The calibration curve covered 3 orders of magnitude with instrument LODs of 1 pg/ μ L injected ($S/N > 3$ for the base peak). An example of quantification is shown in Figure 2 for Mirex and hepta-CB-170. It appeared to be important to carry out the quantification on the entire set of slices corresponding to one analyte. The ratio between corresponding native and ^{13}C -label slices varies over the peak cluster. As illustrated in Table 2, the ratio between native and ^{13}C -label was calculated after the summation of the area of all combined slices.

Human serum QC samples were used to build up QC charts for both GC-IDHRMS and GCxGC-IDTOFMS methods for all investigated analytes. Figure 3 illustrates the case of DBE-47, DBE-100 and CB-153. The SD values are higher with GCxGC-IDTOFMS partly because several slices of different intensities had to be integrated and also due to the low mass resolution of the TOFMS. Finally, OCP (Figure 4) and BDE (Table 3) data are shown for human serum and milk samples, respectively.

Table 1. List of masses used to reconstruct deconvoluted ion currents (DICs) during the GCxGC-IDTOFMS quantification process

	Compounds	^{12}C -native ions (m/z)	^{13}C -labeled ions (m/z)
PCBs	Tri-CBs	186+188+256+258+260	198+200+268+270+272
	Tetra-CBs	220+222+255+257+290+292+294	232+234+267+268+302+304+306
	Penta-CBs	254+256+291+324+326+328	266+268+303+336+338+340
	Hexa-CBs	288+290+292+358+360+362	300+302+304+370+372+374
	Hepta-CBs	322+324+326+394+396+398	334+336+338+406+408+410
	Octa-CBs	356+358+360+362+426+428+430+432	368+370+372+374+438+440+442+444
	Nona-CBs	390+392+394+396+398+460+462+464+466+468	402+404+406+408+410+472+474+476+478+480
	Deca-CBs	424+426+428+430+432+494+496+498+500+502	436+438+440+442+444+506+508+510+512+514
	OCPs		
	HCb	247+249+251+282+284+286	253+255+257+288+290+292
	β - and γ -HCH	217+219+221+252+254+256+288+290+292	223+225+227+258+260+262+294+296+298
	Heptachlor epoxide	351+353+355+357+388+390	361+363+365+367+398+400
	Oxychlorane	385+387+389+391+422+424+426	395+397+399+401+432+434+436
	<i>trans</i> -Nonachlor	407+409+411+413+442+444+446+448	417+419+421+423+452+454+456+458
	Dieldrin	261+263+265+343+345+347+378+380+382	275+277+279+355+357+359+390+390+394
	<i>o,p'</i> - and <i>p,p'</i> -DDT	235+237+239	247+249+251
	Mirex	235+237+239+241+270+272+274+276	240+242+244+246+275+277+279+281
	<i>p,p'</i> -DDE	246+248+250+316+318+320	258+260+262+328+330+332
	PBDEs		
	Tri-BDEs	246+248+404+406+408+410	258+260+416+418+420+422
	Tetra-BDEs	324+326+328+482+484+486+488+490	336+338+340+494+496+498+500+502
	Penta-BDEs	402+404+406+408+562+564+566+568	414+416+418+420+574+576+578+580
	Hexa-BDEs	480+482+484+486+640+642+644+646+648	492+494+496+498+652+654+656+658+660
	Hexa-BB	466+468+470+545+547+549+624+626+628+630	476+478+480+557+559+561+636+638+640+642

ACKNOWLEDGEMENTS

J.-F. Focant was financially supported by the Oak Ridge Institute for Science and Education (ORISE), a Department of Energy (DOE) facility managed by Oak Ridge Associated Universities (ORAU).

REFERENCES

- [1] Sjödin, A.; Jones, R. S.; Lapeza Jr., C. R.; Focant, J.-F.; McGahee III, E. E.; Patterson Jr., D. G. *Anal. Chem.* **2004**, *76*, 1921-1927.
- [2] Sjödin, A.; McGahee III, E.E.; Focant, J.-F.; Jones, R. S., Lapeza Jr., C. R.; Zhang, Y.; Patterson Jr., D. G. *Anal. Chem.* **2004**, in press.
- [3] Sandau, C. D.; Sjödin, A.; Davis, M. D.; Barr, J. R.; Maggio, V. L.; Waterman, A. L.; Preston, K. E.; Preau Jr, J. L.; Barr, D.; Needham, L. L.; Patterson Jr, D. G. *Anal. Chem.* **2003**, *75*, 71-77.
- [4] Focant, J.-F.; Sjödin, A.; Patterson Jr., D. G. *J. Chromatogr. A* **2003**, 1019, 143-156.
- [5] Focant, J.-F.; Sjödin, A.; Patterson Jr., D. G. *J. Chromatogr. A* **2004**, in press.

Table 2. Variation in the native over ^{13}C -label ratio for the different slices of a peak cluster

Compound	1t_R (S)	2t_R (S)	^{13}C -Label		Native		Ratio ^a
			area	S/N	area	S/N	
BDE-99-1	2493	1.72	4755	22.00	786	4.00	0.17
BDE-99-2 ^b	2496	1.68	20523	102.00	4351	20.00	0.21
BDE-99-3 ^c	2499	1.71	18341	78.00	6335	26.00	0.35
BDE-99-4	2502	1.65	3940	21.00	2210	10.00	0.56
Sum	-		47559	-	13682	-	0.29

^aRatio of native over ^{13}C -label. ^bBase peak for ^{13}C -label. ^cBase peak for natives.

Table 3. Comparison between the new GCxGC-IDTOFMS and the reference GC-IDHRMS method for the measurement (ng/g lipids) of selected BDEs in natural human milk pools

Analyte	Pool A					Pool B					Pool C				
	GC-HRMS		GCxGC-TOFMS		Dev. ^b (%)	GC-HRMS		GCxGC-TOFMS		Dev. (%)	GC-HRMS		GCxGC-TOFMS		Dev. (%)
	n=3		n=4			n=3		n=4			n=120		n=4		
	Mean	SEM ^a	Mean	SEM		Mean	SEM	Mean	SEM		Mean	SEM	Mean	SEM	
BDE-28	6.6	0.6	7.7	0.2	15.3	12.7	0.9	11.7	1.0	8.3	2.9	0.1	4.0	0.5	41.0
BDE-47	230.4	8.2	227.4	17.4	1.3	284.5	14.0	308.3	24.4	8.4	64.0	2.8	65.7	0.4	2.6
BDE-100	46.1	1.7	49.4	0.9	7.3	45.6	2.2	53.9	3.7	18.3	11.4	0.6	11.4	0.7	0.3
BDE-99	71.6	2.9	76.7	1.6	7.1	74.0	2.6	81.7	6.4	10.4	19.3	0.7	19.7	0.6	1.9
BB-153	0.7	0.1	1.0	0.3	52.7	1.2	0.0	1.6	0.5	27.7	6.7	0.4	9.0	0.7	33.6
BDE-154	5.9	0.2	4.9	0.4	17.3	3.7	0.2	4.7	1.1	27.3	1.1	0.0	0.9	0.2	18.3
BDE-85	7.9	1.0	6.1	0.3	22.9	6.9	0.2	7.4	0.8	7.1	1.6	0.0	-	-	
BDE-153	18.5	1.4	18.6	0.3	0.7	21.4	0.9	25.8	5.4	20.4	9.2	0.4	9.8	0.9	6.1
ΣBDEs	386.9	36.5	390.8	16.2	1.0	448.8	20.9	493.5	39.0	9.9	109.5	4.7	111.5	2.1	1.8

^a SEM, standard error of the mean. ^b Deviation between the 2 methods.





Figure 1. Contour plot for GCxGC-IDTOFMS TIC chromatogram of a 100 pg/ μ L native compound multi-analytes calibration solution (top) and a real human serum sample (bottom).




Figure 2.

Region of the chromatogram of a real human serum sample (top) where Mirex and CB-170 elute. Extracted ion chromatograms are based on ions listed in Table 1 for those 2 compounds (native and labeled). Peaks are: (1) Mirex, (2) CB-170, (3) phthalate, (4) siloxane bleed, (5) tetracosane. The dashed Gaussian shapes are artificial and are only shown to help to locate the elution windows of the 2 compounds. Expanded section (bottom) showing one of the slices illustrating chromatographic and mass spectral resolution of the 2 compounds and their corresponding ^{13}C labels (left) as well as the corresponding contour plots (right).




Figure 4.

Comparison between GCxGC-IDTOFMS and the reference GC-IDHRMS for the measurement of OCPs in human serum samples.




Figure 3. QC charts for BDE-47 (A), BDE-100 (B), and CB-153 (C).